

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03027 Calib. Date: 02/12/2021 Calib. No.: 00039
Certification File No.: 03028 Cert. Date: 02/12/2021 Cert. No.: 00029
Linearity File No.: 03029 Lin. Date: 02/12/2021 Lin. No.: 00029
Solution File No.: 03056 Soln. Date: 03/18/2021 Soln. No.: 00385
Sequential File No.: 03056 File Date: 03/18/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 0102

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:29S		
Control 1 EC	0.097%	07:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:30S		
Control 2 EC	0.095%	07:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	07:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:32S		
Control 3 EC	0.096%	07:32S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	07:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:33S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  # 350

Badge No.: 350

Date: 03/18/2021