

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03027	Calib. Date: 02/12/2021
Certification File No.:	03028	Calib. No.: 00039
Linearity File No.:	03029	Cert. Date: 02/12/2021
Solution File No.:	03067	Cert. No.: 00029
Sequential File No.:	03067	Lin. Date: 02/12/2021
		Lin. No.: 00029
		Soln. Date: 03/31/2021
		Soln. No.: 00388
		File Date: 03/31/2021

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 09/02/2022
Solution Control Lot:	20390		Bottle No.: 0040

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:16S		
Control 1 EC	0.098%	09:17S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:17S		
Control 2 EC	0.095%	09:18S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19S		
Control 3 EC	0.096%	09:19S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: \_\_\_\_\_

*PH. M. Nater #336*

Badge No.: 336

Date: 03/31/2021