

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03027	Calib. Date: 02/12/2021
Certification File No.:	03028	Calib. No.: 00039
Linearity File No.:	03029	Cert. Date: 02/12/2021
Solution File No.:	03069	Cert. No.: 00029
Sequential File No.:	03069	Lin. Date: 02/12/2021
		Lin. No.: 00029
		Soln. Date: 04/01/2021
		Soln. No.: 00389
		File Date: 04/01/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0699

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:28S		
Control 1 EC	0.099%	12:28S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S		
Control 2 EC	0.098%	12:30S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30S		
Control 3 EC	0.098%	12:31S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 04/01/2021