

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01891	Calib. Date: 02/12/2021
Certification File No.:	01892	Calib. No.: 00051
Linearity File No.:	01893	Cert. Date: 02/12/2021
Solution File No.:	01905	Cert. No.: 00038
Sequential File No.:	01905	Lin. Date: 02/12/2021
		Lin. No.: 00038
		Soln. Date: 04/09/2021
		Soln. No.: 00320
		File Date: 04/09/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0701

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:15D		
Control 1 EC	-.---%	11:16D	23.4°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	11:16D	23.4°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_

*P. P. Gambardella* #365

Badge No.: 365

Date: 04/09/2021