

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03027 Calib. Date: 02/12/2021 Calib. No.: 00039
Certification File No.: 03028 Cert. Date: 02/12/2021 Cert. No.: 00029
Linearity File No.: 03029 Lin. Date: 02/12/2021 Lin. No.: 00029
Solution File No.: 03074 Soln. Date: 04/12/2021 Soln. No.: 00390
Sequential File No.: 03074 File Date: 04/12/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0703

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:50D		
Control 1 EC	0.099%	12:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:51D		
Control 2 EC	0.097%	12:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:52D		
Control 3 EC	0.097%	12:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: _____

PH. M. J. # 336

Badge No.: 336

Date: 04/12/2021