

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03027 Calib. Date: 02/12/2021 Calib. No.: 00039
Certification File No.: 03028 Cert. Date: 02/12/2021 Cert. No.: 00029
Linearity File No.: 03029 Lin. Date: 02/12/2021 Lin. No.: 00029
Solution File No.: 03086 Soln. Date: 04/20/2021 Soln. No.: 00393
Sequential File No.: 03086 File Date: 04/20/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0209

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:05D		
Control 1 EC	0.097%	18:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	18:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:07D		
Control 2 EC	0.095%	18:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	18:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:08D		
Control 3 EC	0.095%	18:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	18:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:09D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature:  #399

Badge No.: 399

Date: 04/20/2021