

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03027	Calib. Date: 02/12/2021
Certification File No.:	03028	Calib. No.: 00039
Linearity File No.:	03029	Cert. Date: 02/12/2021
Solution File No.:	03089	Cert. No.: 00029
Sequential File No.:	03089	Lin. Date: 02/12/2021
		Lin. No.: 00029
		Soln. Date: 04/24/2021
		Soln. No.: 00394
		File Date: 04/24/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0697

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:28D		
Control 1 EC	0.098%	08:29D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:30D		
Control 2 EC	0.096%	08:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31D		
Control 3 EC	0.097%	08:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: _____

Adam Koepen #350

Badge No.: 350

Date: 04/24/2021