

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01891 Calib. Date: 02/12/2021 Calib. No.: 00051
Certification File No.: 01892 Cert. Date: 02/12/2021 Cert. No.: 00038
Linearity File No.: 01893 Lin. Date: 02/12/2021 Lin. No.: 00038
Solution File No.: 01922 Soln. Date: 05/07/2021 Soln. No.: 00323
Sequential File No.: 01922 File Date: 05/07/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0022

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:53D		
Control 1 EC	0.098%	11:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:55D		
Control 2 EC	0.097%	11:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:56D		
Control 3 EC	0.097%	11:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 05/07/2021