

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01891	Calib. Date:	02/12/2021
Certification File No.:	01892	Cert. Date:	02/12/2021
Linearity File No.:	01893	Lin. Date:	02/12/2021
Solution File No.:	01931	Soln. Date:	05/15/2021
Sequential File No.:	01931	File Date:	05/15/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	0727

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:07D		
Control 1 EC	0.099%	23:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:08D		
Control 2 EC	0.097%	23:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:10D		
Control 3 EC	0.097%	23:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:

 #350

Badge No.: 350

Date: 05/15/2021