

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01891	Calib. Date:	02/12/2021
Certification File No.:	01892	Cert. Date:	02/12/2021
Linearity File No.:	01893	Lin. Date:	02/12/2021
Solution File No.:	01941	Soln. Date:	05/28/2021
Sequential File No.:	01941	File Date:	05/28/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	0027

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:17D		
Control 1 EC	0.100%	11:18D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19D		
Control 2 EC	0.098%	11:19D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:20D		
Control 3 EC	0.098%	11:21D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 05/28/2021