

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01891	Calib. Date: 02/12/2021
Certification File No.:	01892	Calib. No.: 00051
Linearity File No.:	01893	Cert. Date: 02/12/2021
Solution File No.:	01944	Cert. No.: 00038
Sequential File No.:	01944	Lin. Date: 02/12/2021
		Lin. No.: 00038
		Soln. Date: 05/31/2021
		Soln. No.: 00326
		File Date: 05/31/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0087

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:42D		
Control 1 EC	0.099%	09:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43D		
Control 2 EC	0.096%	09:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45D		
Control 3 EC	0.096%	09:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTEALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 05/31/2021