

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08753</b>		2. Contact: <b>Chris Fluck</b>		4. Date: <b>04/29/21</b>	
		3. Phone Number: <b>732-349-0150</b>		5. County: <b>Ocean</b>	
6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARAJ-0075</b></p>		7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>		8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	

9. Reason for Service:  
**Complaint received of the instrument experiencing multiple control test failures.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed multiple control test failures due to low electrochemical sensor (EC) results. The instrument was placed out of service and returned to Draeger Inc for evaluation.**

Last Known Sequential File #: 03089

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045  
Name & Badge Number (Print)

Tpr I Lutz 7045 4-29-21  
Signature Date

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08753</b>	2. Contact: <b>Chris Fluck</b> 3. Phone Number: <b>732-349-0150</b>	4. Date: <b>6/23/21</b> 5. County: <b>Ocean</b>
--	--	--

6. Alcotest Instrument Serial Number: <p style="text-align: center;"><b>ARAJ-0075</b></p>	7. Simulator Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;"><b>N/A</b></p>
--	--	--

9. Reason for Service:  
**The alcotest instrument was returned from outside evaluation and placed back in service.**


10. Comments:  
**See Draeger Return and Repair form.**

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE

Last Known Sequential File #: NA

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Dennis J. Lutz #7045 <small>Name &amp; Badge Number (Print)</small>	 <small>Signature</small>	7045 6-23-21 <small>Date</small>
---	--	-------------------------------------

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Township Of Toms River, Toms River, NJ

Date Received: 05/11/2021

Date given to service: 05/11/2021

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR AJ-0075

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	2.0	
MP Labor	Labor	.5	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell and geared engine.

FUEL CELL = ARND-1633

CAL W/QC AND OPS CHECK

Service Technician: CR

Date: 6/15/2021



# Delivery receipt



Customer no.  
150059761

Number of report  
306062626

Please reference on inquiries

Date of receiving / / Delivery receipt date  
06/15/2021

Consignee, 150059763  
**TOWNSHIP OF TOMS RIVER POLICE DEPAR**  
**ATTN: CHRIS FLUCK**  
**PO#ARAJ-0075**  
**255 OAK AVE**  
**TOMS RIVER NJ 08753-3383**

Customer, 150059761  
TOWNSHIP OF TOMS RIVER  
ATTN: CHRIS FLUCK  
PO#ARAJ-0075  
255 OAK AVE  
TOMS RIVER NJ 08753-0876

Your order

Date of order:  
05/03/2021  
Purch.ord.no.:  
21-02855

Branch text

Draeger contact person

Service Support-IDT  
800-437-2437 OPT 5  
SS-IDT@draeger.com

Page 1

Quantity	Description	Part no.	Equipment - No.	Serial - No.
001	ALCOTEST 7110 MK III USA	8314246	1030320499	ARAJ-0075

**Draeger Inc.**  
Our Tax ID: 23-1699096  
3135 Quarry Road, Telford, PA 18969  
An Equal Opportunity Employer M / F / V / H  
Telephone 800-437-2437  
<http://www.draeger.com>

Remit to:  
**LOCKBOX (Standard USPS)**  
Draeger, Inc.  
PO Box 13369  
Newark, New Jersey  
07101-3362

Remit to:  
**LOCKBOX (Overnight)**  
FIS Lockbox Processing  
Lockbox #13369  
100 Grove Road  
Suite E  
West Deptford, NJ 08066

Remit US Wire Transfers to:  
Account Name: Draeger Inc.  
Account Number: 00-494-936  
Transit Routing: 021001033  
SWIFT: BKTRUS33  
Deutsche Bank Trust Company Americas  
60 Wall Street 25th Fl, New York, NY 10005

# Service note

Customer no.                      Number of report                      Date of receiving  
 150059761                      306062626                      05/03/2021  
**Please reference on inquiries**

Customer  
 TOWNSHIP OF TOMS RIVER  
 255 OAK AVE  
 TOMS RIVER NJ 08753-0876

Consignee  
 TOWNSHIP OF TOMS RIVER POLICE DEPAR  
 ATTN: CHIEF OF POLICE  
 255 OAK AVE  
 TOMS RIVER NJ 08753-3383

Your order  
 Date of order:  
 05/03/2021  
 Your reference:  
 21-02855

NL-Text  
 Maintenance device list

Your contact person  
 Service Support-IDT  
 800-437-2437 OPT 5  
 SS-IDT@draeger.com

Page 1

Item	Quantity	Part no.	Description
			<b>Service Order: 144768725</b>
			REPAIR/ARAJ-0075
			Eq.No. : 1030320499    SN: ARAJ-0075
			Material: 8314246 ALCOTEST 7110 MK III USA
			Inv.No. :
			Location:
			<b>REPAIR/ARAJ-0075</b>
	0.5	H	R001 Repair time - normal hours
	1	EA	6808455 Sensor Alcotest B
	1	EA	6808374 Geared Engine, A7110/A9510