

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03094	Calib. Date: 06/23/2021
Certification File No.:	03095	Calib. No.: 00043
Linearity File No.:	03096	Cert. Date: 06/23/2021
Solution File No.:	03097	Cert. No.: 00030
Sequential File No.:	03097	Lin. Date: 06/23/2021
		Lin. No.: 00030
		Soln. Date: 06/23/2021
		Soln. No.: 00396
		File Date: 06/23/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0722

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:09D		
Control 1 EC	0.102%	15:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:11D		
Control 2 EC	0.101%	15:11D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:12D		
Control 3 EC	0.101%	15:12D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-025 04

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I [Signature] 7045

Badge No.: 7045

Date: 06/23/2021