

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03094	Calib. Date: 06/23/2021
Certification File No.:	03095	Calib. No.: 00043
Linearity File No.:	03096	Cert. Date: 06/23/2021
Solution File No.:	03119	Cert. No.: 00030
Sequential File No.:	03119	Lin. Date: 06/23/2021
		Lin. No.: 00030
		Soln. Date: 07/07/2021
		Soln. No.: 00397
		File Date: 07/07/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0698

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:30D		
Control 1 EC	0.099%	04:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	04:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:32D		
Control 2 EC	0.098%	04:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:33D		
Control 3 EC	0.099%	04:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BARTOSHEK

First Name: PTL

MI: D.

Signature: _____

Paul J. Bartoshek #316 PTL

Badge No.: 316

Date: 07/07/2021