

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARAJ-0075 |
| Location: | TOMS RIVER P.D. | |
| Calibration File No.: | 03094 | Calib. Date: 06/23/2021 |
| Certification File No.: | 03095 | Calib. No.: 00043 |
| Linearity File No.: | 03096 | Cert. Date: 06/23/2021 |
| Solution File No.: | 03127 | Cert. No.: 00030 |
| Sequential File No.: | 03127 | Lin. Date: 06/23/2021 |
| | | Lin. No.: 00030 |
| | | Soln. Date: 07/23/2021 |
| | | Soln. No.: 00398 |
| | | File Date: 07/23/2021 |
| | | |
| Calibrating Unit: | WET - | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDWK S3-0386 |
| Solution Control Lot: | 21010 | Expires: 01/13/2023 |
| | | Bottle No.: 0706 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 11:23D | | |
| Control 1 EC | 0.103% | 11:24D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.104% | 11:24D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:24D | | |
| Control 2 EC | 0.102% | 11:25D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.103% | 11:25D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:25D | | |
| Control 3 EC | 0.102% | 11:26D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.102% | 11:26D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:27D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: *P. P. [Signature]* #365

Badge No.: 365

Date: 07/23/2021