

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03094 Calib. Date: 06/23/2021 Calib. No.: 00043
Certification File No.: 03095 Cert. Date: 06/23/2021 Cert. No.: 00030
Linearity File No.: 03096 Lin. Date: 06/23/2021 Lin. No.: 00030
Solution File No.: 03136 Soln. Date: 07/28/2021 Soln. No.: 00399
Sequential File No.: 03136 File Date: 07/28/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0708

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:19D		
Control 1 EC	0.099%	12:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:20D		
Control 2 EC	0.097%	12:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21D		
Control 3 EC	0.098%	12:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 07/28/2021