

Calibrating Unit*

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01975	Calib. Date: 08/06/2021
Certification File No.:	01976	Calib. No.: 00052
Linearity File No.:	01977	Cert. Date: 08/06/2021
Solution File No.:	01978	Cert. No.: 00039
Sequential File No.:	01978	Lin. Date: 08/06/2021
		Lin. No.: 00039
		Soln. Date: 08/06/2021
		Soln. No.: 00330
		File Date: 08/06/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	21010	Expires: 01/13/2023
		Bottle No.: 0707

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:05D		
Control 1 EC	0.101%	15:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06D		
Control 2 EC	0.100%	15:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		
Control 3 EC	0.099%	15:08D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-026 (DU)

Changed By:
 Last Name: LUTZ First Name: DENNIS MI: J
 Signature: Tpr I Dent 7045 Badge No.: 7045
 Date: 08/06/2021