

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03094	Calib. Date: 06/23/2021
Certification File No.:	03095	Calib. No.: 00043
Linearity File No.:	03096	Cert. Date: 06/23/2021
Solution File No.:	03148	Cert. No.: 00030
Sequential File No.:	03148	Lin. Date: 06/23/2021
		Lin. No.: 00030
		Soln. Date: 08/10/2021
		Soln. No.: 00400
		File Date: 08/10/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21150	Expires: 05/04/2023
		Bottle No.: 0679

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:24D		
Control 1 EC	0.096%	12:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		
Control 2 EC	0.096%	12:26D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26D		
Control 3 EC	0.096%	12:27D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHWARTZ

First Name: STEVE

MI: -

Signature: _____



Badge No.: 1293

Date: 08/10/2021