

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01975 Calib. Date: 08/06/2021 Calib. No.: 00052  
Certification File No.: 01976 Cert. Date: 08/06/2021 Cert. No.: 00039  
Linearity File No.: 01977 Lin. Date: 08/06/2021 Lin. No.: 00039  
Solution File No.: 01993 Soln. Date: 08/22/2021 Soln. No.: 00331  
Sequential File No.: 01993 File Date: 08/22/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387  
Control Solution %: 0.100% Expires: 05/04/2023  
Solution Control Lot: 21150 Bottle No.: 0680

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:54D		
Control 1 EC	0.097%	08:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		
Control 2 EC	0.095%	08:56D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56D		
Control 3 EC	0.097%	08:57D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature: RF #399

Badge No.: 1399

Date: 08/22/2021