

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. No.:	00043
Calibration File No.:	03094	Cert. Date:	06/23/2021
Certification File No.:	03095	Lin. Date:	06/23/2021
Linearity File No.:	03096	Soln. Date:	09/03/2021
Solution File No.:	03178	File Date:	09/03/2021
Sequential File No.:	03178		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	0089

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:23D		
Control 1 EC	0.101%	11:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		
Control 2 EC	0.099%	11:25D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		
Control 3 EC	0.099%	11:26D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_

*P. P. Gambardella #365*

Badge No.: 365

Date: 09/03/2021