

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01975	Calib. Date:	08/06/2021
Certification File No.:	01976	Calib. No.:	00052
Linearity File No.:	01977	Cert. Date:	08/06/2021
Solution File No.:	01996	Cert. No.:	00039
Sequential File No.:	01996	Lin. Date:	08/06/2021
		Lin. No.:	00039
		Soln. Date:	09/10/2021
		Soln. No.:	00332
		File Date:	09/10/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	1401

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:41D		
Control 1 EC	0.100%	11:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		
Control 2 EC	0.098%	11:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43D		
Control 3 EC	0.098%	11:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA	First Name: PASCAL	MI: M
Signature: <u></u>	Badge No.: 365	Date: 09/10/2021