

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01975 Calib. Date: 08/06/2021 Calib. No.: 00052
Certification File No.: 01976 Cert. Date: 08/06/2021 Cert. No.: 00039
Linearity File No.: 01977 Lin. Date: 08/06/2021 Lin. No.: 00039
Solution File No.: 01999 Soln. Date: 09/17/2021 Soln. No.: 00333
Sequential File No.: 01999 File Date: 09/17/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 05/04/2023
Solution Control Lot: 21150 Bottle No.: 0033

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:45D		
Control 1 EC	0.100%	11:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		
Control 2 EC	0.098%	11:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		
Control 3 EC	0.098%	11:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:49D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 09/17/2021