

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03094	Calib. Date: 06/23/2021
Certification File No.:	03095	Calib. No.: 00043
Linearity File No.:	03096	Cert. Date: 06/23/2021
Solution File No.:	03189	Cert. No.: 00030
Sequential File No.:	03189	Lin. Date: 06/23/2021
		Lin. No.: 00030
		Soln. Date: 10/01/2021
		Soln. No.: 00406
		File Date: 10/01/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21150	Expires: 05/04/2023
		Bottle No.: 0149

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:13D		
Control 1 EC	0.099%	12:14D	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:14D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		
Control 2 EC	0.097%	12:15D	33.8°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:15D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16D		
Control 3 EC	0.097%	12:16D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:16D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:17D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 10/01/2021