

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03094	Calib. Date: 06/23/2021
Certification File No.:	03095	Calib. No.: 00043
Linearity File No.:	03096	Cert. Date: 06/23/2021
Solution File No.:	03193	Cert. No.: 00030
Sequential File No.:	03193	Lin. Date: 06/23/2021
		Lin. No.: 00030
		Soln. Date: 10/10/2021
		Soln. No.: 00407
		File Date: 10/10/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	21150	Expires: 05/04/2023
		Bottle No.: 0550

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:16D		
Control 1 EC	0.097%	17:16D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	17:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:17D		
Control 2 EC	0.095%	17:18D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:18D		
Control 3 EC	0.096%	17:19D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFADL

First Name: ROBERT

MI: S

Signature: \_\_\_\_\_



Badge No.: 306

Date: 10/10/2021