

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	03094	Calib. Date:	06/23/2021
Certification File No.:	03095	Calib. No.:	00043
Linearity File No.:	03096	Cert. Date:	06/23/2021
Solution File No.:	03213	Cert. No.:	00030
Sequential File No.:	03213	Lin. Date:	06/23/2021
		Lin. No.:	00030
		Soln. Date:	11/05/2021
		Soln. No.:	00410
		File Date:	11/05/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	0498

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:44S		
Control 1 EC	0.098%	22:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45S		
Control 2 EC	0.097%	22:45S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:46S		
Control 3 EC	0.097%	22:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:47S		

All tests within acceptable tolerance.

On this date, I installed the above-indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

[Handwritten Signature] #365

Badge No.: 365

Date: 11/05/2021