

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01975	Calib. Date: 08/06/2021
Certification File No.:	01976	Calib. No.: 00052
Linearity File No.:	01977	Cert. Date: 08/06/2021
Solution File No.:	02024	Cert. No.: 00039
Sequential File No.:	02024	Lin. Date: 08/06/2021
		Lin. No.: 00039
		Soln. Date: 11/12/2021
		Soln. No.: 00336
		File Date: 11/12/2021

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWK S3-0387
Control Solution %: 0.100%		Expires: 05/04/2023
Solution Control Lot: 21150		Bottle No.: 0087

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:26S		
Control 1 EC	0.100%	11:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27S		
Control 2 EC	0.097%	11:28S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29S		
Control 3 EC	0.097%	11:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 11/12/2021