

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	03094	Calib. Date:	06/23/2021
Certification File No.:	03095	Cert. Date:	06/23/2021
Linearity File No.:	03096	Lin. Date:	06/23/2021
Solution File No.:	03225	Soln. Date:	11/19/2021
Sequential File No.:	03225	File Date:	11/19/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	0211

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:27S		
Control 1 EC	0.099%	12:28S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:28S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28S		
Control 2 EC	0.098%	12:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30S		
Control 3 EC	0.098%	12:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 11/19/2021