

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01975	Calib. Date: 08/06/2021
Certification File No.:	01976	Calib. No.: 00052
Linearity File No.:	01977	Cert. Date: 08/06/2021
Solution File No.:	02028	Cert. No.: 00039
Sequential File No.:	02028	Lin. Date: 08/06/2021
		Lin. No.: 00039
		Soln. Date: 12/01/2021
		Soln. No.: 00337
		File Date: 12/01/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	21150	Expires: 05/04/2023
		Bottle No.: 0437

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:58S		
Control 1 EC	0.099%	10:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00S		
Control 2 EC	0.096%	11:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01S		
Control 3 EC	0.098%	11:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Pascal Gambardella #365

Badge No.: 365

Date: 12/01/2021