

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	06/23/2021
Calibration File No.:	03094	Calib. No.:	00043
Certification File No.:	03095	Cert. Date:	06/23/2021
Linearity File No.:	03096	Lin. Date:	06/23/2021
Solution File No.:	03238	Soln. Date:	12/10/2021
Sequential File No.:	03238	File Date:	12/10/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWK S3-0386
Control Solution %:	0.100%			Expires:	05/04/2023
Solution Control Lot:	21150			Bottle No.:	0212

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:03S		
Control 1 EC	0.099%	13:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:04S		
Control 2 EC	0.097%	13:04S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:05S		
Control 3 EC	0.098%	13:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: SCHWARTZ

First Name: STEVE

MI: -

Signature: \_\_\_\_\_



Badge No.: 1293

Date: 12/10/2021