

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	02034	Calib. Date:	12/16/2021
Certification File No.:	02035	Calib. No.:	00053
Linearity File No.:	02036	Cert. Date:	12/16/2021
Solution File No.:	02037	Cert. No.:	00040
Sequential File No.:	02037	Lin. Date:	12/16/2021
		Lin. No.:	00040
		Soln. Date:	12/16/2021
		Soln. No.:	00338
		File Date:	12/16/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	21150	Expires:	05/04/2023
		Bottle No.:	0677

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:45S		
Control 1 EC	0.101%	13:46S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 2 EC	0.098%	13:47S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		
Control 3 EC	0.098%	13:49S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-444 DL

**Changed By:**

Last Name: LUTZ

First Name: DENNIS

MI: J

Signature: Tpr I [Signature] 7045

Badge No.: 7045

Date: 12/16/2021