Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.				
Calibration File No.:	03241	Calib. Date:	12/16/2021	Calib. No.:	00044
Certification File No.:	03242	Cert. Date:	12/16/2021	Cert. No.:	00031
Linearity File No.:	03243	Lin. Date:	12/16/2021	Lin. No.:	00031
Solution File No.:	03244	Soln. Date:	12/16/2021	Soln. No.:	00414
Sequential File No.:	03244	File Date:	12/16/2021		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAN-0041
Control Solution %:	0.100%			Expires:	05/04/2023
Solution Control Lot:	21150			Bottle No.:	0145

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:22S		
Control 1 EC	0.100%	11:23S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23S		
Control 2 EC	0.098%	11:24S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25S		
Control 3 EC	0.098%	11:25S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-436 (DC)

Changed By:

Last Name: LUTZ First Name: DENNIS MI: J

Signature: Tpr I Dent 7045 Badge No.: 7045

Date: 12/16/2021