

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 03241
Certification File No.: 03242
Linearity File No.: 03243
Solution File No.: 03266
Sequential File No.: 03266
Calib. Date: 12/16/2021
Cert. Date: 12/16/2021
Lin. Date: 12/16/2021
Soln. Date: 01/25/2022
File Date: 01/25/2022
Serial No.: ARAJ-0075
Calib. No.: 00044
Cert. No.: 00031
Lin. No.: 00031
Soln. No.: 00418
Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 21270
Model No.: CU-34
Serial No.: DDAN-0041
Expires: 07/21/2023
Bottle No.: 0945

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:55S		
Control 1 EC	0.100%	14:56S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57S		
Control 2 EC	0.098%	14:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58S		
Control 3 EC	0.097%	14:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHWARTZ

First Name: STEVE

MI: -

Signature: _____

Badge No.: 1293

Date: 01/25/2022