

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	02034	Calib. Date: 12/16/2021
Certification File No.:	02035	Calib. No.: 00053
Linearity File No.:	02036	Cert. Date: 12/16/2021
Solution File No.:	02042	Cert. No.: 00040
Sequential File No.:	02042	Lin. Date: 12/16/2021
		Lin. No.: 00040
		Soln. Date: 01/27/2022
		Soln. No.: 00340
		File Date: 01/27/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 1094

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:04S		
Control 1 EC	0.101%	23:04S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:05S		
Control 2 EC	0.099%	23:06S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:07S		
Control 3 EC	0.099%	23:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

P. P. G. #365

Badge No.: 365

Date: 01/27/2022