

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03241 **Calib. Date:** 12/16/2021 **Calib. No.:** 00044
Certification File No.: 03242 **Cert. Date:** 12/16/2021 **Cert. No.:** 00031
Linearity File No.: 03243 **Lin. Date:** 12/16/2021 **Lin. No.:** 00031
Solution File No.: 03284 **Soln. Date:** 02/14/2022 **Soln. No.:** 00422
Sequential File No.: 03284 **File Date:** 02/14/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAN-0041
Control Solution %: 0.100% **Expires:** 07/21/2023
Solution Control Lot: 21270 **Bottle No.:** 1268

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:06S		
Control 1 EC	0.098%	10:06S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07S		
Control 2 EC	0.095%	10:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.095%	10:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09S		
Control 3 EC	0.095%	10:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CASSADY

First Name: MICHAEL

MI: P

Signature: _____



Badge No.: 367

Date: 02/14/2022