

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03241	Calib. Date: 12/16/2021
Certification File No.:	03242	Calib. No.: 00044
Linearity File No.:	03243	Cert. Date: 12/16/2021
Solution File No.:	03287	Cert. No.: 00031
Sequential File No.:	03287	Lin. Date: 12/16/2021
		Lih. No.: 00031
		Soln. Date: 02/21/2022
		Soln. No.: 00423
		File Date: 02/21/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0980

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:02S		
Control 1 EC	0.100%	15:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03S		
Control 2 EC	0.098%	15:04S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05S		
Control 3 EC	0.097%	15:05S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: _____

[Handwritten Signature] #336

Badge No.: 336

Date: 02/21/2022