

# Calibrating Unit New Standard Solution Report

|                         |                       |              |            |
|-------------------------|-----------------------|--------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARAJ-0075  |
| Location:               | TOMS RIVER P.D.       |              |            |
| Calibration File No.:   | 03241                 | Calib. Date: | 12/16/2021 |
| Certification File No.: | 03242                 | Cert. No.:   | 00044      |
| Linearity File No.:     | 03243                 | Lin. Date:   | 12/16/2021 |
| Solution File No.:      | 03290                 | Soln. No.:   | 00031      |
| Sequential File No.:    | 03290                 | File Date:   | 02/25/2022 |
|                         |                       |              |            |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34      |
| Control Solution %:     | 0.100%                | Serial No.:  | DDAN-0041  |
| Solution Control Lot:   | 21270                 | Expires:     | 07/21/2023 |
|                         |                       | Bottle No.:  | 0934       |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 23:12S        |                               |                           |
| Control 1 EC      | 0.098%         | 23:12S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 23:12S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 23:13S        |                               |                           |
| Control 2 EC      | 0.097%         | 23:14S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 23:14S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 23:14S        |                               |                           |
| Control 3 EC      | 0.097%         | 23:15S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 23:15S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 23:16S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:



Badge No.: 365

Date: 02/25/2022