


Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03241	Calib. Date: 12/16/2021
Certification File No.:	03242	Calib. No.: 00044
Linearity File No.:	03243	Cert. Date: 12/16/2021
Solution File No.:	03292	Cert. No.: 00031
Sequential File No.:	03292	Lin. Date: 12/16/2021
		Lin. No.: 00031
		Soln. Date: 02/26/2022
		Soln. No.: 00425
		File Date: 02/26/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0949

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:14S		
Control 1 EC	0.098%	02:14S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:15S		
Control 2 EC	0.096%	02:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16S		
Control 3 EC	0.097%	02:17S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: GAMBARDELLA First Name: PASCAL MI: M
 Signature:  Badge No.: 365
 Date: 02/26/2022