

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	02034	Calib. Date: 12/16/2021
Certification File No.:	02035	Calib. No.: 00053
Linearity File No.:	02036	Cert. Date: 12/16/2021
Solution File No.:	02058	Cert. No.: 00040
Sequential File No.:	02058	Lin. Date: 12/16/2021
		Lin. No.: 00040
		Soln. Date: 02/28/2022
		Soln. No.: 00343
		File Date: 02/28/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0952

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:29S		
Control 1 EC	0.101%	21:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:30S		
Control 2 EC	0.098%	21:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32S		
Control 3 EC	0.097%	21:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BARTOSHEK

First Name: DAVID

MI: G

Signature:

David M Bartoshek #316 TRPD

Badge No.: 316

Date: 02/28/2022