

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03241	Calib. Date: 12/16/2021
Certification File No.:	03242	Calib. No.: 00044
Linearity File No.:	03243	Cert. Date: 12/16/2021
Solution File No.:	03310	Cert. No.: 00031
Sequential File No.:	03310	Lin. Date: 12/16/2021
		Lin. No.: 00031
		Soln. Date: 03/19/2022
		Soln. No.: 00428
		File Date: 03/19/2022

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAN-0041
Control Solution %:	0.100%			Expires:	07/21/2023
Solution Control Lot:	21270			Bottle No.:	0931

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:06S		
Control 1 EC	0.099%	22:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:07S		
Control 2 EC	0.096%	22:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:09S		
Control 3 EC	0.097%	22:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:10S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 03/19/2022