

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARAJ-0075 |
| Location: | TOMS RIVER P.D. | Calib. Date: | 12/16/2021 |
| Calibration File No.: | 03241 | Calib. No.: | 00044 |
| Certification File No.: | 03242 | Cert. Date: | 12/16/2021 |
| Linearity File No.: | 03243 | Lin. Date: | 12/16/2021 |
| Solution File No.: | 03312 | Soln. Date: | 03/20/2022 |
| Sequential File No.: | 03312 | File Date: | 03/20/2022 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDAN-0041 |
| Control Solution %: | 0.100% | | | Expires: | 07/21/2023 |
| Solution Control Lot: | 21270 | | | Bottle No.: | 0856 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 16:33S | | |
| Control 1 EC | 0.099% | 16:33S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 16:33S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:34S | | |
| Control 2 EC | 0.096% | 16:35S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 16:35S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:36S | | |
| Control 3 EC | 0.097% | 16:36S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 16:36S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 16:37S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 03/20/2022