

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03241	Calib. No.: 00044
Certification File No.:	03242	Cert. No.: 00031
Linearity File No.:	03243	Lin. No.: 00031
Solution File No.:	03322	Soln. No.: 00431
Sequential File No.:	03322	File Date: 04/03/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	21270	Expires: 07/21/2023
		Bottle No.: 0775

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:06D		
Control 1 EC	0.099%	10:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07D		
Control 2 EC	0.097%	10:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		
Control 3 EC	0.098%	10:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: \_\_\_\_\_

*PH. M. Nater #336*

Badge No.: 336

Date: 04/03/2022