

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	12/16/2021
Calibration File No.:	03241	Calib. No.:	00044
Certification File No.:	03242	Cert. Date:	12/16/2021
Linearity File No.:	03243	Lin. Date:	12/16/2021
Solution File No.:	03331	Soln. Date:	04/09/2022
Sequential File No.:	03331	File Date:	04/09/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	0663

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:30D		
Control 1 EC	0.100%	22:31D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:31D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:31D		
Control 2 EC	0.097%	22:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:33D		
Control 3 EC	0.098%	22:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SIENKIEWICZ

First Name: TPR

MI: -

Signature: _____

TPR [Signature] #7933

Badge No.: 7933

Date: 04/09/2022