

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	12/16/2021
Calibration File No.:	03241	Calib. No.:	00044
Certification File No.:	03242	Cert. Date:	12/16/2021
Linearity File No.:	03243	Lin. Date:	12/16/2021
Solution File No.:	03333	Soln. Date:	04/10/2022
Sequential File No.:	03333	File Date:	04/10/2022

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAN-0041
Control Solution %:	0.100%			Expires:	07/21/2023
Solution Control Lot:	21270			Bottle No.:	0887

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:48D		
Control 1 EC	0.099%	14:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:50D		
Control 2 EC	0.097%	14:50D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51D		
Control 3 EC	0.097%	14:52D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: FITZGERALD

First Name: RYAN

MI: P

Signature: _____

[Handwritten Signature] #399

Badge No.: 1399

Date: 04/10/2022