

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 03241 Calib. Date: 12/16/2021 Calib. No.: 00044
Certification File No.: 03242 Cert. Date: 12/16/2021 Cert. No.: 00031
Linearity File No.: 03243 Lin. Date: 12/16/2021 Lin. No.: 00031
Solution File No.: 03340 Soln. Date: 04/15/2022 Soln. No.: 00434
Sequential File No.: 03340 File Date: 04/15/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 12/28/2023
Solution Control Lot: 21480 Bottle No.: 0659

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:50D		
Control 1 EC	0.098%	10:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51D		
Control 2 EC	0.096%	10:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52D		
Control 3 EC	0.097%	10:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SCHWARTZ

First Name: STEVE

MI: -

Signature: _____



Badge No.: 1293

Date: 04/15/2022