

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03241	Calib. Date: 12/16/2021
Certification File No.:	03242	Calib. No.: 00044
Linearity File No.:	03243	Cert. Date: 12/16/2021
Solution File No.:	03353	Cert. No.: 00031
Sequential File No.:	03353	Lin. Date: 12/16/2021
		Lin. No.: 00031
		Soln. Date: 04/30/2022
		Soln. No.: 00437
		File Date: 04/30/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 0685

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:21D		
Control 1 EC	0.099%	03:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:23D		
Control 2 EC	0.097%	03:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:24D		
Control 3 EC	0.097%	03:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	03:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 04/30/2022