

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03241	Calib. Date: 12/16/2021
Certification File No.:	03242	Calib. No.: 00044
Linearity File No.:	03243	Cert. Date: 12/16/2021
Solution File No.:	03358	Cert. No.: 00031
Sequential File No.:	03358	Lin. Date: 12/16/2021
		Lin. No.: 00031
		Soln. Date: 05/09/2022
		Soln. No.: 00438
		File Date: 05/09/2022

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 12/28/2023
Solution Control Lot:	21480		Bottle No.: 0388

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:40D		
Control 1 EC	0.100%	15:40D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:41D		
Control 2 EC	0.098%	15:42D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:43D		
Control 3 EC	0.099%	15:43D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____



Badge No.: 306

Date: 05/09/2022