

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	03241	Calib. Date: 12/16/2021
Certification File No.:	03242	Calib. No.: 00044
Linearity File No.:	03243	Cert. Date: 12/16/2021
Solution File No.:	03360	Cert. No.: 00031
Sequential File No.:	03360	Lin. Date: 12/16/2021
		Lin. No.: 00031
		Soln. Date: 05/12/2022
		Soln. No.: 00439
		File Date: 05/12/2022

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAN-0041
Control Solution %:	0.100%		Expires: 12/28/2023
Solution Control Lot:	21480		Bottle No.: 0560

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:31D		
Control 1 EC	0.100%	08:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:33D		
Control 2 EC	0.098%	08:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34D		
Control 3 EC	0.098%	08:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature: \_\_\_\_\_

*Adam Koepen* #350

Badge No.: 350

Date: 05/12/2022