

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 02034 Calib. Date: 12/16/2021 Calib. No.: 00053  
Certification File No.: 02035 Cert. Date: 12/16/2021 Cert. No.: 00040  
Linearity File No.: 02036 Lin. Date: 12/16/2021 Lin. No.: 00040  
Solution File No.: 02087 Soln. Date: 05/27/2022 Soln. No.: 00348  
Sequential File No.: 02087 File Date: 05/27/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040  
Control Solution %: 0.100% Expires: 12/28/2023  
Solution Control Lot: 21480 Bottle No.: 0386

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:27D		
Control 1 EC	0.104%	08:28D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:28D		
Control 2 EC	0.098%	08:29D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.094%	08:29D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	08:30D		
Control 3 EC	0.099%	08:30D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.093%	08:30D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	08:31D		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #1350

Badge No.: 350

Date: 05/27/2022