

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	02034	Calib. Date: 12/16/2021
Certification File No.:	02035	Calib. No.: 00053
Linearity File No.:	02036	Cert. Date: 12/16/2021
Solution File No.:	02097	Cert. No.: 00040
Sequential File No.:	02097	Lin. Date: 12/16/2021
		Lin. No.: 00040
		Soln. Date: 06/07/2022
		Soln. No.: 00350
		File Date: 06/07/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0040
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 0630

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:57D		
Control 1 EC	0.099%	11:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58D		
Control 2 EC	0.096%	11:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		
Control 3 EC	0.098%	12:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NATER

First Name: MARK

MI: S

Signature: _____



Badge No.: 336

Date: 06/07/2022