

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 03241 Calib. Date: 12/16/2021 Calib. No.: 00044  
Certification File No.: 03242 Cert. Date: 12/16/2021 Cert. No.: 00031  
Linearity File No.: 03243 Lin. Date: 12/16/2021 Lin. No.: 00031  
Solution File No.: 03374 Soln. Date: 06/07/2022 Soln. No.: 00441  
Sequential File No.: 03374 File Date: 06/07/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041  
Control Solution %: 0.100% Expires: 12/28/2023  
Solution Control Lot: 21480 Bottle No.: 0579

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:49D		
Control 1 EC	0.097%	11:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50D		
Control 2 EC	0.095%	11:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		
Control 3 EC	0.096%	11:52D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: \_\_\_\_\_

*PH. NATER* # 336

Badge No.: 336

Date: 06/07/2022